

PROCEEDINGS OF SPIE

SPIDigitalLibrary.org/conference-proceedings-of-spie

Linear-array based full-view high-resolution photoacoustic computed tomography of whole mouse brain functions in vivo

Lei Li, Pengfei Zhang, Lihong V. Wang

Lei Li, Pengfei Zhang, Lihong V. Wang, "Linear-array based full-view high-resolution photoacoustic computed tomography of whole mouse brain functions in vivo," Proc. SPIE 10494, Photons Plus Ultrasound: Imaging and Sensing 2018, 104941H (19 February 2018); doi: 10.1117/12.2289118

SPIE.

Event: SPIE BiOS, 2018, San Francisco, California, United States

Linear-array based full-view high-resolution photoacoustic computed tomography of whole mouse brain functions *in vivo*

Lei Li^{a†}, Pengfei Zhang^{b†}, Lihong V. Wang^{a*}

^aCaltech Optical Imaging Laboratory, Andrew and Peggy Cherng Department of Medical Engineering and Department of Electrical Engineering, California Institute of Technology, 1200 E California Blvd., Pasadena, CA 91125; ^bDepartment of Biomedical Engineering, Washington University in St. Louis, One Brookings Dr., St. Louis, MO, 63130

*LVW@Caltech.edu; †The authors contribute equally to the manuscript.

ABSTRACT

Photoacoustic computed tomography (PACT) is a non-invasive imaging technique offering high contrast, high resolution, and deep penetration in biological tissues. We report a photoacoustic computed tomography (PACT) system equipped with a high frequency linear array for anatomical and functional imaging of the mouse whole brain. The linear array was rotationally scanned in the coronal plane to achieve the full-view coverage. We investigated spontaneous neural activities in the deep brain by monitoring the hemodynamics and observed strong interhemispherical correlations between contralateral regions, both in the cortical layer and in the deep regions.

Keywords: Photoacoustic computed tomography; deep brain imaging; functional imaging

1. INTRODUCTION

In recent years, advanced imaging methodologies with increased spatial and temporal resolution have been used to extend our understanding of brain functions. Brain positron emission tomography (PET) is a metabolic imaging method that detects the emissions from radioactive tracers that are injected into the blood stream and eventually reach the brain [1]. Functional magnetic resonance imaging (fMRI) measures brain activity by using the blood oxygenation level-dependent (BOLD) effect [2]. Both methods have drawbacks: PET uses ionizing radiation and has poor spatial resolution, and fMRI is too slow for studying fast hemodynamics. Recently, functional ultrasound imaging has successfully been used to monitor cerebral microvascular hemodynamics deep in the rat brain with high spatiotemporal resolution [3, 4]. Optical imaging techniques, such as optical coherence tomography and multiphoton microscopy, have also been demonstrated for imaging neural activity with cellular and subcellular resolution [5, 6], but the shallow penetration limits their observation to only the cortical layer (~1–2 mm).

Photoacoustic tomography (PAT) is a modality that detects ultrasonic waves generated by optical absorption of soft tissues under pulsed light illumination [7, 8]. PAT is a nonionizing imaging technique, providing strong optical contrast and high ultrasonic spatial resolution at depths. By using multiple wavelength lasers to excite photoacoustic waves, PAT is able to simultaneously measure the concentrations of various chromophores in biological tissues or organs with distinct optical absorption properties. In particular, PAT has been successfully used to quantify the concentration and the oxygen saturation (sO_2) of hemoglobin, which are related to angiogenesis and hypermetabolism [9, 10]. Substantially accurate sO_2 quantification has also been demonstrated in deep tissues [11]. By scanning a tightly focused laser beam, a PAT system can provide optical resolution photoacoustic microscopy (OR-PAM) [12–14]. Recently, OR-PAM has been used for high-speed, high-resolution mapping of the blood network and for studying the hemodynamics in the mouse brain [15–17]. Because OR-PAM relies on ballistic photons to achieve high resolution, its penetration depth in tissue is generally limited to ~1–2 mm. Photoacoustic computed tomography (PACT) detects ultrasonic waves generated by diffused photons and retrieves the absorption distribution through an inverse algorithm, allowing an imaging depth of up to several centimeters [18, 19]. In this work, we used PACT to image the blood vascular network of a whole mouse brain through the intact skull, with high resolution and substantial depth. We then used the PACT system to study the hemodynamics of the mouse brain in the resting state. Our experiments show that functional PACT is a promising tool for studying neural activities in the deep brain, and potentially for revealing the mechanism of some brain disorders.

2. METHODS

Figure 1 shows the schematic setup of linear array PACT for mouse brain imaging. In this work, 6–8 week-old female mice (Swiss Webster, Invigo) were used. Prior to imaging, the hair on the head of the mouse was removed by a depilatory cream. The mouse was then mounted onto a holder with its head fixed by a nose cone and a tooth bar. A rubber tube connecting the nose cone and an isoflurane vaporizer was used to deliver oxygen and anesthetic gas. The temperature of the mouse holder was regulated by a controller. The scalp of the mouse was removed, and ultrasound gel was applied before imaging. The animal procedures followed the laboratory animal protocols approved by the Animal Studies Committee of Washington University in St. Louis. The mouse was placed beneath the water tank, within an imaging window with dimensions of $8\text{ cm} \times 10\text{ cm}$. The imaging window was covered with plastic film at the bottom of the water tank, and the tank was filled with water for ultrasound coupling. A 5 cm wide window in one side of the tank wall allowed laser beam access to the mouse head. The laser beam used for exciting ultrasonic waves was provided by a Nd:YAG laser (Quantel, Brilliant B, 4–6 ns pulse duration, 10 Hz repetition rate). The laser beam was expanded and homogenized by an engineered diffuser (EDC-5-A-2 s, RPC Photonics, Inc.), resulting in a 2 cm-diameter illumination area on the mouse head. With a pulse energy of 200 mJ, the optical fluence on the head surface was 64 mJ/cm^2 , which is below the American National Standards Institute (ANSI) safety limit of 100 mJ/cm^2 at 1064 nm. The excited photoacoustic (PA) waves were detected by a focused linear transducer array (LZ250, VisualSonics) consisting of 256 elements. The linear array had a one-way bandwidth of 11 MHz around the central frequency of 21 MHz, providing a lateral resolution of 75 microns. The numerical aperture of the transducer array is 0.1, yielding a sectioning thickness of about 0.5 mm within the depth of focus, which covers the whole mouse brain. The detected signals were first amplified by homemade pre-amplifiers (256 channels, 26 dB Gain per channel) and then digitalized by a customized 256-channel data acquisition system (National Instruments, Customized PXI system, 14 bits) at a sampling rate of 250 MHz. The features within the imaging plane were reconstructed using the filtered universal back-projection algorithm [20].




Fig. 1 Schematic setup of linear array PACT for mouse brain imaging. Pre-Amp, pre-amplifier; DAQ, data acquisition unit, PC, personal computer

To map the vessels in the coronal plane, the mouse was mounted with its cortical plane horizontal to the bottom of the water tank, and the linear array was oriented in the coronal plane, as shown in Fig. 1. The surface of the linear array was about 1 cm away from the top of the head, and the diffused laser beam was obliquely incident on the head. The pulsed laser at 1064 nm was used to excite the PA waves, which came predominantly from oxyhemoglobin, a dominant chromophore in the blood stream. Due to the limited view of the linear array, only vessels normal to the acoustic axis could be reconstructed, but the vessels with their orientation deviated from the axis were almost invisible [21].

3. RESULTS

Fig. 2(a-c) show reconstructed images of a typical coronal plane recorded from several representative view angles, each averaged over 200 acquisitions. To solve this limited-view problem, the linear array was mounted on a rotary stage (not shown in Fig. 1) with its rotational axis centered at the mouse brain. The linear array was then scanned around the brain without changing the laser illumination to record images at view angles ranging from -76 to 76 degrees, with a step size of 4 degrees. The reconstructed bipolar images were converted into unipolar images by Hilbert transformation along each view's acoustic axis [22]. The processed images were then rotated back to recover the vascular distribution in the global coordinate system in which the brain was fixed, and the rotated images were summed to form a full-view unipolar image, as shown in Fig. 2(d). The full-view image demonstrated that the PACT system equipped with the high frequency linear array can provide high resolution and deep penetration imaging of a whole mouse brain.




Fig. 2 Coronal plane PACT imaging of a mouse brain through the intact skull (Bregma -1.0 mm). (a-c) Limited-view planar images reconstructed at view angles of -76 , 0 , 76 degrees, respectively. Each of these images was averaged over 200 acquisitions. (d) Full-view planar image combined from the 39 limited-view images. The scale bar is 2 mm.

Since light at the chosen laser wavelength is weakly absorbed in biological tissues, and changes in the concentrations of hemoglobin and other absorbers during neural activities are generally small, the local optical fluence in the mouse brain is assumed to be temporally invariant, enabling the PACT system to study the hemodynamic activities in the deep brain. As a first demonstration, we used our linear-array PACT system to study the resting-state functional connectivity (RSFC) in the coronal plane of a mouse brain. RSFC measures the temporal correlation of the low frequency, spontaneous hemodynamic fluctuations among spatially separated but functionally related regions in the brain [23, 24]. The alteration in the functional connectivity properties is usually an indicator of such brain disorders as depression, Alzheimer's disease, and schizophrenia [25-27]. To explore the functional connectivity of a mouse brain in resting state, the linear array was fixed with its axis normal to the cortical plane, and 6000 images were acquired at a frame rate of 10 Hz. This study again used 6–8 week-old female ND4 Swiss Webster mice (Invigo), anesthetized with 0.8% (vol/vol) isoflurane at an air flow rate of 0.8 L/min, a dose does not alter the functional connectivity patterns [28]. The PA signals due to scattered photons from the transducer surface were used to correct the fluctuations in the total laser pulse energy. Every set of 10 images was averaged to improve the signal-to-noise (SNR) ratio, and the resulting 600 images over a 10-

minute acquisition were analyzed with a seed-based approach [29, 30]. Basically, the reconstructed images were first filtered by a 2-D Gaussian smoothing kernel with standard deviation of 6 pixels. Then a band pass filter with a window of 0.009-0.08 Hz was applied on the time series of the images to reduce the effects caused by heart beating and breathing [30]. Finally, a seed was chosen and the temporal correlations of its PA signal with that of each pixel in the image were calculated. A high coefficient in the correlation map indicates a region that is functionally related to the seed region. The correlation maps for representative regions in the mouse brain were superimposed on the synthesized anatomic image obtained as described above. As shown in Fig. 3a, b, strong interhemispherical correlations were observed in the chosen functional regions, including the somatosensory, retrohippocampal regions. The high correlations in the contralateral hemisphere were mostly from the regions symmetrically opposed to the seeds. Figure 3(c) also shows the time series of the filtered PA signals of a pair of correlated regions marked in white and black circles in Fig. 3(a), displaying strong correlation in their fluctuations.




Fig. 3 Functional connectivity in a mouse brain acquired by a linear-array PACT system. (a, b) The correlation maps calculated at representative functional regions (black circles) are overlaid onto the anatomic image that was synthesized from multiple limited-view images (Bregma -1.0 mm). The scale bar is 2 mm. S1HL, somatosensory 1; RHP, retrohippocampal region. (c) Time series of the filtered PA signals of a pair of correlated regions as marked in (a). The red and blue curves are corresponding to the points marked in white and black circles, respectively. Amplitudes are normalized to the maximum.

4. DISCUSSION

In summary, we have built a linear-array PACT system for imaging a whole mouse brain through the intact skull and for studying hemodynamic activities with high spatial resolution and deep penetration. The linear array was scanned in the coronal plane to collect images at different angles. Combining the limited-view images into a full-view image revealed the microvascular network in the brain. We investigated spontaneous neural activities in the deep brain by monitoring the concentration of hemoglobin in the blood vessels and observed strong interhemispherical correlations between chosen functional regions, both in the cortex and in the deep region.

The imaging frame rate of our functional PACT system is lower than that of functional ultrasound imaging [3, 4], but this limitation can be circumvented by employing a higher repetition-rate laser for photoacoustic excitation. However, PACT utilizes a different contrast mechanism from ultrasound imaging. Ultrasound tomography images the mechanical properties of soft tissue, and it has poor extravascular molecular contrasts. On the basis of the optical absorption of biomolecules, PACT enables molecular imaging by matching the excitation optical wavelength with the absorption peaks of the targeted molecules [31, 32], and thus provides high image contrast. Moreover, the concentration and oxygen saturation of hemoglobin, which cannot be obtained by ultrasound imaging but are closely related to brain activities, can be measured by PACT [11].

This work was funded by the National Institutes of Health (NIH) Grants U01 NS090579 (NIH Brain Initiative Award), U01 NS099717 (NIH Brain Initiative Award), R01 EB016963, DP1 EB016986 (Director's Pioneer Award), R01 CA186567 (Director's Transformative Research Award), S10 RR026922. The authors thank Prof. James Ballard for his close reading of the manuscript. L.V. Wang has a financial interest in Microphotoacoustics, Inc., CalPACT, LLC, and Union Photoacoustic Technologies, Ltd., which, however, did not support this work.

REFERENCES

- [1] L. L. Firestone, F. Gyulai, M. Mintun *et al.*, "Human brain activity response to fentanyl imaged by positron emission tomography," *Anesthesia & Analgesia*, 82(6), 1247-1251 (1996).
- [2] S.-G. Kim, and S. Ogawa, "Biophysical and physiological origins of blood oxygenation level-dependent fMRI signals," *Journal of Cerebral Blood Flow & Metabolism*, 32(7), 1188-1206 (2012).
- [3] B.-F. Osmanski, S. Pezet, A. Ricobaraza *et al.*, "Functional ultrasound imaging of intrinsic connectivity in the living rat brain with high spatiotemporal resolution," *Nature communications*, 5, 5023 (2014).
- [4] E. Macé, G. Montaldo, I. Cohen *et al.*, "Functional ultrasound imaging of the brain," *Nature methods*, 8(8), 662 (2011).
- [5] Y. Chen, A. D. Aguirre, L. Ruvinskaya *et al.*, "Optical coherence tomography (OCT) reveals depth-resolved dynamics during functional brain activation," *Journal of neuroscience methods*, 178(1), 162-173 (2009).
- [6] C. Tischbirek, A. Birkner, H. Jia *et al.*, "Deep two-photon brain imaging with a red-shifted fluorometric Ca²⁺ indicator," *Proceedings of the National Academy of Sciences*, 112(36), 11377-11382 (2015).
- [7] L. Li, J. Yao, L. V. Wang *et al.*, [Photoacoustic Tomography Enhanced by Nanoparticles] John Wiley & Sons, Inc., (1999).
- [8] M. Zhou, L. Li, J. Yao *et al.*, [Nanoparticles for Photoacoustic Imaging of Vasculature] Springer International Publishing, Cham(2017).
- [9] L. Lin, J. Yao, L. Li *et al.*, "In vivo photoacoustic tomography of myoglobin oxygen saturation." 21, 5.
- [10] L. Li, L. Zhu, C. Ma *et al.*, "Single-impulse panoramic photoacoustic computed tomography of small-animal whole-body dynamics at high spatiotemporal resolution," *Nature Biomedical Engineering*, 1, 0071 (2017).
- [11] S. Tzoumas, A. Nunes, I. Olefir *et al.*, "Eigenspectra optoacoustic tomography achieves quantitative blood oxygenation imaging deep in tissues," *Nature communications*, 7, ncomms12121 (2016).
- [12] L. Li, C. Yeh, S. Hu *et al.*, "Fully motorized optical-resolution photoacoustic microscopy," *Optics Letters*, 39(7), 2117-2120 (2014).
- [13] Y. S. Zhang, J. Yao, C. Zhang *et al.*, "Optical-Resolution Photoacoustic Microscopy for Volumetric and Spectral Analysis of Histological and Immunochemical Samples," *Angewandte Chemie International Edition*, 53(31), 8099-8103 (2014).
- [14] L. Zhu, L. Li, L. Gao *et al.*, "Multiview optical resolution photoacoustic microscopy," *Optica*, 1(4), 217-222 (2014).
- [15] J. Yao, L. Wang, J.-M. Yang *et al.*, "High-speed label-free functional photoacoustic microscopy of mouse brain in action," *Nature Methods*, 12, 407 (2015).
- [16] B. Rao, R. Zhang, L. Li *et al.*, "Photoacoustic imaging of voltage responses beyond the optical diffusion limit," *Scientific Reports*, 7(1), 2560 (2017).
- [17] W. Liu, Y. Zhou, M. Wang *et al.*, "Correcting the limited view in optical-resolution photoacoustic microscopy," *Journal of Biophotonics*, e201700196-n/a.
- [18] L. Li, J. Xia, G. Li *et al.*, "Label-free photoacoustic tomography of whole mouse brain structures *ex vivo*." 3, 8.
- [19] C. Yeh, L. Li, L. Zhu *et al.*, "Dry coupling for whole-body small-animal photoacoustic computed tomography." 22, 5.
- [20] M. Xu, and L. V. Wang, "Universal back-projection algorithm for photoacoustic computed tomography," *Physical Review E*, 71(1), 016706 (2005).
- [21] Y. Xu, L. V. Wang, G. Ambartsoumian *et al.*, "Reconstructions in limited - view thermoacoustic tomography," *Medical physics*, 31(4), 724-733 (2004).
- [22] G. Li, L. Li, L. Zhu *et al.*, "Multiview Hilbert transformation for full-view photoacoustic computed tomography using a linear array," *Journal of biomedical optics*, 20(6), 066010 (2015).

- [23] B. Biswal, F. Zerrin Yetkin, V. M. Haughton *et al.*, "Functional connectivity in the motor cortex of resting human brain using echo - planar mri," *Magnetic resonance in medicine*, 34(4), 537-541 (1995).
- [24] M. D. Fox, and M. E. Raichle, "Spontaneous fluctuations in brain activity observed with functional magnetic resonance imaging," *Nature reviews neuroscience*, 8(9), 700 (2007).
- [25] M. D. Greicius, B. H. Flores, V. Menon *et al.*, "Resting-state functional connectivity in major depression: abnormally increased contributions from subgenual cingulate cortex and thalamus," *Biological psychiatry*, 62(5), 429-437 (2007).
- [26] R. L. Buckner, J. Sepulcre, T. Talukdar *et al.*, "Cortical hubs revealed by intrinsic functional connectivity: mapping, assessment of stability, and relation to Alzheimer's disease," *Journal of neuroscience*, 29(6), 1860-1873 (2009).
- [27] S. M. Lawrie, C. Buechel, H. C. Whalley *et al.*, "Reduced frontotemporal functional connectivity in schizophrenia associated with auditory hallucinations," *Biological psychiatry*, 51(12), 1008-1011 (2002).
- [28] J. Grandjean, A. Schroeter, I. Batata *et al.*, "Optimization of anesthesia protocol for resting-state fMRI in mice based on differential effects of anesthetics on functional connectivity patterns," *Neuroimage*, 102, 838-847 (2014).
- [29] M. D. Fox, D. Zhang, A. Z. Snyder *et al.*, "The global signal and observed anticorrelated resting state brain networks," *Journal of neurophysiology*, 101(6), 3270-3283 (2009).
- [30] M. Nasirivanaki, J. Xia, H. Wan *et al.*, "High-resolution photoacoustic tomography of resting-state functional connectivity in the mouse brain," *Proceedings of the National Academy of Sciences*, 111(1), 21-26 (2014).
- [31] A. P. Jathoul, J. Laufer, O. Ogunlade *et al.*, "Deep in vivo photoacoustic imaging of mammalian tissues using a tyrosinase-based genetic reporter," *Nature Photonics*, 9(4), 239 (2015).
- [32] J. Yao, A. A. Kaberniuk, L. Li *et al.*, "Multiscale photoacoustic tomography using reversibly switchable bacterial phytochrome as a near-infrared photochromic probe," *Nature methods*, 13(1), 67 (2016).